

CERTIFICATE OF PERFORMANCE

On this date: July 2 2013 The Cleanroom facilities of:

Mountain Integrated Medical Devices

Has been tested in accordance with the International Standard ISO 14644-1

in the unoccupied state, and meets:

ISO Class 7 at: 0.5, 1.0, and 5.0 micron particulate sizes



Tested by: Controlled Environment Technologies
299 Knottwood Road North Edmonton AB Canada T6K 2G5

Performance Technician: Chris Steinke Signature

A handwritten signature in black ink, appearing to read 'Chris', written over a horizontal line.